

Peracetic Acid (PAA) Test Kit

KS-0099B

1 Drop = 13 ppm PAA

1. Fill the graduated reaction vial (PS-1060) to 10 mL with the water sample.
2. Add 10 drops of KI solution (RS-0005) and mix by swirling.
3. Add 10 drops of Phosphoric Acid Reagent (RS-0012) and mix by swirling.
4. Add 5 drops of Starch Indicator Solution (RS-0008) and mix by swirling. The solution will turn brown.
5. Titrate with Sodium Thiosulfate 0.10 N (RS-0002), drop by drop, mixing after each drop until the sample turns colorless passing through blue.

ppm PAA = # of Drops of Titrating Solution Used to Change the Color x 13

Replacement Reagents & Parts

RS-0002	2oz	Sodium Thiosulfate 0.10 N Solution
RS-0005	2oz	Potassium Iodide (KI) Solution
RS-0012	2oz	Phosphoric Acid Reagent
RS-0008	2oz	Starch Indicator Solution
PS-1060	Each	Graduated Reaction Vial
PW-1110	Each	Test Kit Box w/ Foam Insert & Bottom Foam Pad